

CERTIFICATE OF ANALYSIS

Page 1 of 1

CUSTOMER
Addmaster Ltd
Darfin House
Priestly Court
Staffordshire Technology Park
Stafford
ST18 0AR
UK.

CERTIFICATE NO. 1026632.207/9457
CUSTOMER REF. 2324

SAMPLE DETAILS DATE RECEIVED 10/02/2015
MEDISHOWER ORDER NO.

METHOD: Determination of Antibacterial Activity using Test Based on MOD ISO 22196

DATE ANALYSED 11/02/2015 DATE REPORTED 13/02/2015

RESULTS (AS CFU CM²)

SAMPLE	SPECIES	CONTACT TIME		REDUCTION (INITIAL)	
		0 hrs	24 hrs	Log 10	%
POLYPROPYLENE	MRSA	1.6E+04	8.9E+03		
BIOMASTER TREATED MEDISHOWER (BIOMASTER 612 @2%)	MRSA	1.6E+04	< 25.00	≥ 2.81	≥ 99.84%

Key: NS = Poor survival on control supplied.

The above data show the difference in the population following contact with the surface of the samples listed for 24 hours at 35°C under a RH of > 95% relative to the initial population.

IMSL MICROBIOLOGICAL SERVICES LTD
PALE LANE
HARTLEY WINTNEY
HANTS RG27 8DH
UK

MANAGING DIRECTOR
Peter D Askew

